

Adjustment Report



Part No (P/N) 226130
Serial No (S/N) 85791671

DMA 35 Ampere

Serial Numbers

Ser. No. Main Board	41391631		
Ser. No. Cartridge	39973771		
Ser. No. Front	17447	Date	29.07.2026

Calibration

	Temperature [°C]	Density [g/cm³]	Result*
Deionized water	13,4	0,9993	Passed
Deionized water	21,1	0,9979	Passed
Deionized water	37,6	0,9931	Passed
Air	21,8	0,0012	Passed
Ethanol (96% v/v)	21,6	0,8063	Passed

* Density Tolerance $\pm 0.001 \text{ g/cm}^3$

Firmware

Version	1.016
---------	-------

Anton Paar GmbH
Anton-Paar-Str. 20
8054 Graz, Austria

Inspector: KKR2125
Inspection date: 10.08.2026
Form: E28AF19-B

Certificate of Compliance

Part No (P/N) 226130
Serial No (S/N) 85791671

DMA 35 Ampere

Dear customer,

With this certificate the manufacturer declares that this instrument was inspected, tested and accepted in accordance with our quality-management system. It complies with the specifications according to the part number listed above. The following tests have been successfully performed:

- ☒ Optical check of electronic boards, wiring system and surface finish.
- ☒ Functional test of the instrument (display, touch-buttons, hand pump, automatic shut-down and acoustic signal)
- ☒ Test of interfaces.
- ☒ Measurements: Adjustment with air at 20 °C and distilled water at 13 °C, 20 °C and 40 °C.
- ☒ Check with distilled water and Alcohol at different temperatures.
- ☒ Check of all labels
- ☒ Final check

Anton Paar GmbH
Anton-Paar-Str. 20
8054 Graz, Austria

Inspector: KKR2125
Inspection date: 10.08.2026
Form: E28AF20-A